

초저온 냉동고 온도 기록지(Ⅱ)

장비명	Deep Freezer		
모델명	ULT-837		
Serial No	GS28190109-12		
설정온도	-80 ℃		
위 치	임상시험센터		
날짜 (YYYY.MM.DD)	min ℃	max ℃	기타사항(알람 등)
2026-04-01	-81.11	-76.87	
2026-04-02	-81.12	-76.85	
2026-04-03	-81.11	-76.88	
2026-04-04	-81.10	-76.86	
2026-04-05	-81.12	-76.88	
2026-04-06	-81.11	-76.87	
2026-04-07	-81.12	-76.40	
2026-04-08	-81.15	-76.12	
2026-04-09	-81.12	-76.84	
2026-04-10	-81.13	-76.92	
2026-04-11	-81.10	-76.85	
2026-04-12	-81.08	-76.27	
2026-04-13	-81.11	-76.12	
2026-04-14	-81.12	-76.82	
2026-04-15	-81.15	-76.10	
2026-04-16	-81.16	-76.91	
2026-04-17	-81.17	-76.90	
2026-04-18	-81.16	-76.97	
2026-04-19	-81.18	-76.84	
2026-04-20	-81.16	-76.13	
2026-04-21	-81.14	-76.91	
2026-04-22	-81.12	-76.89	
2026-04-23	-81.11	-76.85	
2026-04-24	-81.16	-76.85	
2026-04-25	-81.09	-76.84	
2026-04-26	-81.12	-76.82	
2026-04-27	-81.13	-76.84	
2026-04-28	-81.14	-76.45	
2026-04-29	-81.16	-76.89	
2026-04-30	-81.14	-76.88	
점검자: <u>박선영</u> 이름/서명 점검일: <u>2026.05.06</u> YYYY.MM.DD			