



정기점검 보고서

제품정보: DISCOVERY DMI 20CM 4 RING

서비스 번호: 13768227

시스템 ID: 0850450200

제조번호: COL201800044PT

계약번호: KOR_Contract_SA-2240247

의료기기 품목명 : 표시기재사항 참고 / 의료기기 수입자명 : 지이헬스케어코리아
의료기기 제조국 : 표시기재사항 참고 / 제조연월일 : 표시기재사항 참고

고객정보			
병원명:	이화여자대학교의과대학부속서울병원		
병원주소:	서울특별시 강서구 공항대로 260 이화의대부속서울병원 SEOUL Korea, Republic of 157-210		
담당선생님:	--	Equipment Number:	--
연락처:	--	Room/Department:	0850450200
Email Address:	--	구매번호:	NO PO

Field Service No. WO-15227624 on 2025-12-26				
PM Level of Service: C				
PM Sequence: 5				
고장 내용:	PM on Dec.			
수리/점검내역:	System backup was done. No error was shown in GFV Time & Date was normal PARC Health viewer was normal Chiller coolant was refilled DAS filter was cleaned Slipring brush was cleaned Slipring cover was cleaned PET calibration was done - Dectector cal & CTC cal PET DQA was done Tube warmup was done CT-QA was done			
테스트 내용:	Test scan was successful.			
작업				
작업 종류	엔지니어	시작	종료	기간
Travel	junseong Son	13:00	14:00	1h 0min
정기점검	junseong Son	14:00	18:00	4h 0min
Travel	junseong Son	18:00	19:00	1h 0min
Tool Information: --				

엔지니어 서명:

엔지니어 이름: junseong Son

날짜/시간: 2025-12-30 13:18



15th Floor, Seoul Square, 416
Hangang-daero,
Jung-gu, Seoul 04637 Korea
Customer Service Center 1544-6119

정기점검 보고서

병원명: 핵의학과
Customer Designation:
Customer Document ID:

저희진
↳ ㅇ ↳