

검체 냉장고 온도 기록지(Ⅱ)

장비명	Refrigerator		
모델명	CBR-150-182		
Serial No	GS24190115-24		
설정온도	4 °C		
위 치	임상시험센터		
날짜 (YYYY.MM.DD)	min °C	max °C	기타사항(알람 등)
2026-06-01	3.56	4.41	
2026-06-02	3.58	4.35	
2026-06-03	3.57	4.55	
2026-06-04	3.56	4.57	
2026-06-05	3.56	4.35	
2026-06-06	3.59	4.33	
2026-06-07	3.54	4.31	
2026-06-08	3.56	4.57	
2026-06-09	3.56	4.41	
2026-06-10	3.54	4.51	
2026-06-11	3.57	4.49	
2026-06-12	3.55	4.54	
2026-06-13	3.59	4.50	
2026-06-14	3.56	4.47	
2026-06-15	3.55	4.52	
2026-06-16	3.55	4.49	
2026-06-17	3.53	4.50	
2026-06-18	3.54	4.41	
2026-06-19	3.54	4.56	
2026-06-20	3.56	4.47	
2026-06-21	3.54	4.57	
2026-06-22	3.57	4.55	
2026-06-23	3.53	4.57	
2026-06-24	3.55	4.62	
2026-06-25	3.51	4.47	
2026-06-26	3.53	4.61	
2026-06-27	3.54	4.60	
2026-06-28	3.55	4.58	
2026-06-29	3.54	4.49	
2026-06-30	3.53	4.51	

점검자 : 박신영 /
 이름/서명

점검일 : 2026.07.01
 YYYY.MM.DD