

검체 냉장고 온도 기록지(Ⅱ)

장비명	Refrigerator		
모델명	CBR-150-182		
Serial No	GS24190115-24		
설정온도	4 ℃		
위 치	임상시험센터		
날짜 (YYYY.MM.DD)	min ℃	max ℃	기타사항(알람 등)
2026-07-01	3.55	4.53	
2026-07-02	3.57	4.52	
2026-07-03	3.56	4.67	
2026-07-04	3.55	4.57	
2026-07-05	3.56	4.55	
2026-07-06	3.55	4.49	
2026-07-07	3.57	4.55	
2026-07-08	3.56	4.93	
2026-07-09	3.55	4.51	
2026-07-10	3.55	4.49	
2026-07-11	3.54	4.44	
2026-07-12	3.55	4.56	
2026-07-13	3.54	4.52	
2026-07-14	3.53	4.48	
2026-07-15	3.54	4.59	
2026-07-16	3.55	4.58	
2026-07-17	3.54	4.54	
2026-07-18	3.56	4.53	
2026-07-19	3.53	4.56	
2026-07-20	3.55	4.53	
2026-07-21	3.54	4.47	
2026-07-22	3.56	4.48	
2026-07-23	3.58	4.46	
2026-07-24	3.57	4.65	
2026-07-25	3.55	4.40	
2026-07-26	3.56	4.48	
2026-07-27	3.55	4.55	
2026-07-28	3.56	4.65	
2026-07-29	3.58	4.48	
2026-07-30	3.55	4.54	
2026-07-31	3.53	4.58	
점검자: <u>백선영 / m</u> 이름/서명 점검일: <u>2026.08.03</u> YYYY.MM.DD			