

초저온 냉동고 온도 기록지(Ⅱ)

장비명	Deep Freezer		
모델명	ULT-837		
Serial No	GS28190109-12		
설정온도	-80 ℃		
위 치	임상시험센터		
날짜 (YYYY.MM.DD)	min ℃	max ℃	기타사항(알람 등)
2026-07-01	-81.12	-76.67	
2026-07-02	-81.11	-75.20	
2026-07-03	-81.15	-76.93	
2026-07-04	-81.13	-76.92	
2026-07-05	-81.14	-76.93	
2026-07-06	-81.13	-76.95	
2026-07-07	-81.20	-76.93	
2026-07-08	-81.11	-76.91	
2026-07-09	-81.13	-76.92	
2026-07-10	-81.10	-76.91	
2026-07-11	-81.13	-76.92	
2026-07-12	-81.12	-76.93	
2026-07-13	-81.13	-76.95	
2026-07-14	-81.14	-76.94	
2026-07-15	-81.13	-76.95	
2026-07-16	-81.12	-76.95	
2026-07-17	-81.13	-76.95	
2026-07-18	-81.10	-76.90	
2026-07-19	-81.13	-76.92	
2026-07-20	-81.11	-76.92	
2026-07-21	-81.11	-76.91	
2026-07-22	-81.10	-76.93	
2026-07-23	-81.09	-76.95	
2026-07-24	-81.07	-76.88	
2026-07-25	-81.07	-76.92	
2026-07-26	-81.13	-76.89	
2026-07-27	-81.12	-76.58	
2026-07-28	-81.11	-76.91	
2026-07-29	-81.15	-76.95	
2026-07-30	-81.12	-76.98	
2026-07-31	-81.08	-76.91	
점검자 : <u>백선영 / </u> 이름/서명 점검일 : <u>2026. 08. 03</u> YYYY.MM.DD			